


诊断用固定阳极X射线管 | 执行标准 Standard : YZB/浙 0326 浙食药监械(准)字200第2310012号
管子型号TUBE TYPE | XD7-1.05/35

主要用途


APPLICATION



主要供医学诊断X射线机配套使用
It is designed for main applications used in x-ray equipment for medical diagnosis

外形与接线图

OUTLINE & WIRING DRAWING




技术参数

TECHNICAL DATA

标称X射线管电压 nominal X-ray tube voltage (kV)	靶 target		灯丝特性 filament characteristic	焦点标称值 nominal focal spot value
	靶材 material	靶面角 angle(°)		
35	Mo	19	$I_f=4A$ 时, $U_f=3.8V\pm 0.4V$	1.0

阳极标称输入功率 nominal anode input power (kW) (1s)	固有滤过 inherent filtration (mmAl)	质量 weight (kg)
1.05	≤1.0	≤0.7

灯丝特性曲线 FILAMENT CHARACTERISTIC



阴极发射特性曲线 CATHODE EMISSION CHARACTERISTIC

